



DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-19480
Mfr. Reference: GER00365

Centers for Disease Control
and Prevention (CDC)
National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
P.O. Box 18070
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
July 25, 2014

Mr. Geoff White
Louis M. Gerson Company, Inc.
16 Commerce Blvd.
Middleboro, MA 02346

Dear Mr. White:


The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted January 30, 2014. This request was for approval of the model 1735 filtering facepiece air purifying respirator for protections against particulates at a N95 filter efficiency level, reference assembly matrix prbb51B30AM35.xls, Revision B35 dated January 15, 2014. In addition, private label versions of the model 1735 filtering facepiece respirator were added for Libus of Buenos Aires, Argentina [REDACTED] as the same model number and are identified on the same assembly matrix prbb51B30AM35.xls. Revision B35 dated January 15, 2014.

This request is granted. Approvals are granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-84A-6952 has been assigned. These respirators are approved for protection against particulates at a N95 filter efficiency level.

The CD enclosed with this letter contains the final respirator approval labels. The abbreviated labels have been accepted as submitted. The cautions and limitations which apply to this approval are on the approval labels. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assemblies consist of the parts as listed on the approval labels and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).


This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).



The manufacturer is responsible for properly packaging, labeling, and controlling the respirators produced under these private label approvals. At a minimum, the items that must be controlled are the approved user instructions, all approval labeling, all approved packaging, use claims, marketing materials, and the respirator design and construction details. Any change to these NIOSH-approved respirators or approval documentation without prior notification and approval is a violation of this approval and renders this certification as invalid.

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely yours,



Jonathan V. Szalajda
Acting Chief, Technology Evaluation Branch
National Personal Protective Technology Laboratory

Enclosures

